



**ДЕРЖАВНА СЛУЖБА З ЛІКАРСЬКИХ ЗАСОБІВ ТА КОНТРОЛЮ ЗА
НАРКОТИКАМИ У КИЇВСЬКІЙ ОБЛАСТІ**

пров. Світличної Надії, 3, м.Київ, 02099, тел/факс: (044) 576-40-41
E-mail: dls.ko@dls.gov.ua, Код ЄДРПОУ 37078774

**ВИСНОВОК
про якість ввезеного в Україну лікарського засобу**

15.01.2025

№ 668/25/10

ЛІПРИМАР®

(найменування лікарського засобу згідно з реєстраційним посвідченням)

**таблетки, вкриті плівковою оболонкою, по 20 мг, по 10 таблеток у блістері; по 3
блістери у картонній коробці**

(форма випуску, дозування, вид пакування лікарського засобу)

Номер реєстраційного посвідчення UA/2377/01/01 строк дії реєстраційного посвідчення 01.01.2099

Серія лікарського засобу № **LG4956**

Кількість ввезеного лікарського засобу 720

Виробник

Пфайзер Менюфектуринг Дойчленд ГмбХ, Німеччина

(найменування виробника лікарського засобу, країна походження)

Ввезено в Україну

**Спільне українсько-естонське підприємство у формі товариства
з обмеженою відповідальністю "Оптіма-Фарм, ЛТД", ідент. код:
21642228**

(найменування та код за ЄДРПОУ юридичної особи або прізвище, ім'я, по батькові фізичної
особи - підприємця, її місце проживання та реєстраційний номер облікової картки платника
податків або серія та номер паспорта)

Протокол візуального контролю від 13.01.2025 № 0037/6.

За результатами державного контролю встановлено, що лікарський засіб ввезено в Україну з
дотриманням вимог законодавства щодо забезпечення якості лікарських засобів.

Заступник начальника Служби
(посадова особа органу державного контролю)



(підпис)

Ольга ЄРЬОМЕНКО

(ініціали та прізвище)

	ПФАЙЗЕР МЕНЮФЕКЧУРИНГ ДОЙЧЛЕНД ГМБХ МУСВАЛЬДАЛЬ 1 79108 ФРАЙБУРГ ІМ БРАЙСГАУ НІМЕЧЧИНА	дата 02-09-2024 Сформовано: Наталія Вилензек Сторінка 1 з 4
---	---	--

Сертифікат аналізу**Ліпримар®, таблетки, вкриті плівковою оболонкою, по 20 мг****№10х3 у блістерах**

Призначення: Україна

 Номер матеріалу: F000142816
 Дата виробництва: 15/11/2023

 Серія №: LG4956
 Термін придатності: 01/11/2026
 Аналітична процедура: 966000-09

Тест	Метод	Допустимі межі	Результат
ОПИС Результат	і 2.02/ візуально	Білі круглі, покриті плівковою оболонкою таблетки з гравіруванням «20» з одного боку та «ATV» - з іншого.	відповідає
РОЗЧИНЕННЯ АТОРВАСТАТИНУ через 15 хвилин			
Результат	TM-0258A/УФ	Не менше 80% (Q) від заявленої кількості	відповідає
- середнє значення	TM-0258A/УФ	-	92
- мінімум	TM-0258A/УФ	-	87
- максимум	TM-0258A/УФ	-	100
- стандартне відхилення	TM-0258A/УФ	-	4,7
- стадія	TM-0258A/УФ	-	1
ІДЕНТИФІКАЦІЯ АТОРВАСТАТИНУ / ТШХ			
Результат	TM-0253A/ТШХ	Величина R _f для основної плями на хроматограмі, одержаній для випробовуваного розчину, повинна відповідати величині R _f для основної плями на хроматограмі, одержаній для стандартного розчину	відповідає
ІДЕНТИФІКАЦІЯ АТОРВАСТАТИНУ / ВЕРХ			
Результат	TM-0255A/ВЕРХ	Час утримування основного піка, одержаного з хроматограми розчину для кількісного визначення, повинен відповідати часу утримування основного піка, одержаного із хроматограми стандартного розчину	відповідає
ЧИСТОТА АТОРВАСТАТИНУ / ВЕРХ			
Індивідуальні відомі Продукти розпаду PD140728 Результат	TM-0255A/ВЕРХ	не більше ніж 0.3% (площа)	0.2
- PD130694 Результат	TM-0255A/ВЕРХ	не більше ніж 0.3% (площа)	<0.1
- PD153824 Результат	TM-0255A/ВЕРХ	не більше ніж 0.3% (площа)	<=0.3
- PD148996 Результат	TM-0255A/ВЕРХ	не більше ніж 0.3% (площа)	<=0.3
- PD139884 Результат	TM-0255A/ВЕРХ	не більше ніж 0.3% (площа)	<=0.3

Документ розглядається як власність та був створений для операцій бізнесу та для перегляду робітниками та представниками регуляторних органів.
 Розповсюдження та передача третім сторонам без дозволу заборонено.

Handwritten signature:
 08.09.2024

	ПФАЙЗЕР МЕНЮФЕКЧУРИНГ ДОЙЧЛЕНД ГМБХ МУСВАЛЬДАЛЬ 1 79108 ФРАЙБУРГ ІМ БРАЙСГАУ НІМЕЧЧИНА	дата 02-09-2024 Сформовано: Наталія Вилензек Сторінка 2 з 4
---	---	---

Сертифікат аналізу

Ліпримар®, таблетки, вкриті плівковою оболонкою, по 20 мг №10х3 у блістерах

Призначення - Україна

Номер матеріалу: F000142816
Дата виробництва: 15/11/2023

Серія №: LG4956
Термін придатності: 01/11/2026
Аналітична процедура: 966000-09

Тест	Метод	Допустимі межі	Результат
Індивідуальні невідомі продукти розпаду чи невідомі домішки окремо Результат	TM-0255A/ВЕРХ	не більше ніж 0.2% (площа)	<=0.2
Загальна кількість продуктів розпаду Результат	TM-0255A/ВЕРХ	не більше ніж 1.5% (площа)	0.2
КІЛЬКІСНИЙ ВМІСТ АТОРВАСТАТИНУ /ВЕРХ Результат	TM-0255A/ ВЕРХ	19.4 – 21.0 мг (заяв. 20 мг)	20,4
КІЛЬКІСНИЙ ВМІСТ АТОРВАСТАТИНУ /ВЕРХ Результат	TM-0255A/ ВЕРХ	97.0 - 105.0 %	102.2
ОДНОРІДНІСТЬ ДОЗОВАНИХ ОДИНИЦЬ/УФ Результат		відповідає вимогам Євр.Фарм. 2.9.40	відповідає
- середнє значення	TM-0256A/ Євр.Фарм. 2.9.40/USP-NF	-	101.7
- мінімум	<905>	-	100.1
- максимум		-	103.2
- стадія		-	1
- критерій прийнятності		не більше ніж 15 %	3
- коефіцієнт відхилення		-	1.0

Сертифікат аналізу був створений валідованою системою після погодження керівником відділу контролю якості і дійсний без підпису.

Написання адреси та поштового індексу виробника ПФайзер Менюфекчуринг Дойчленд ГмбХ було змінено та раніше відомі як
Пфайзер Менюфекчуринг Дойчленд ГмбХ, БЕТРІБШТЕТТЕ ФРАЙБУРГ, МУСВАЛЬДАЛЬ 1, 79090 ФРАЙБУРГ

Виправлення поштового індексу та написання адреси виробника ПФайзер Менюфекчуринг Дойчленд ГмбХ не впливає на місцезнаходження виробничої дільниці та не впливає на вміст виробничої ліцензії та сертифікату НВП.

Документ розглядається як власність, та був створений для операцій бізнесу та для перегляду робітниками та представниками регуляторних органів.
Розповсюдження та передача третім сторонам без дозволу заборонено.

	ПФАЙЗЕР МЕНЮФЕКЧУРИНГ ДОЙЧЛЕНД ГМБХ МУСВАЛЬДАЛЬ 1 79108 ФРАЙБУРГ ІМ БРАЙСГАУ НІМЕЧЧИНА	дата 02-09-2024 Сформовано. Наталія Вилензек Сторінка 3 з 4
---	---	--

Сертифікат відповідності

Ліпримар®, таблетки, вкриті плівковою оболонкою, по 20 мг №10х3 у блістерах

Призначення - Україна -

Матеріал №:	F000142816
Серія №:	LG4956
Дата виробництва:	15/11/2023
Термін придатності:	01/11/2026
Вихід:	22544 уп.
Дата випуску:	13/06/2024
Вихідні серії:	3193512
Реєстраційне посвідчення №:	UA/2377/01/01
Назва діючої речовини:	аторвастатин кальцію
Тип упаковки:	блістер
Лікарська форма:	таблетки, вкриті плівковою оболонкою

Назви, адреси та номери ліцензій для всіх дільниць по виробництву та контролю якості лікарського засобу:
 Пфайзер Менюфекчуринг Дойчленд ГмбХ, Мусвальдаль 1, 79108 Фрайбург Ім Брайсгау, Німеччина
 DE_BW_01_MIA_2024_0007

Мікробіологічна чистота:

-загальний вміст аеробних бактерій, метод Євр.Фарм. 2.6.12, границі – не більше 10^3 КУО/г;

-загальний вміст дріжджів та плісневих грибів, метод Євр.Фарм. 2.6.12, границі – не більше 10^2 КУО/г;

специфічні мікроорганізми, E.coli – метод Євр.Фарм. 2.6.13, границі - відповідають.

Це тестування не проводиться на кожній серії; відповідає вимогам при тестуванні. Якщо в сертифікаті аналізу не вказані результати по даному тесту, це означає, що тестування не проводилось.

Сертифікат якості

Органам Охорони Здоров'я України

ГАРАНТІЙНИЙ ЛИСТ

Я, той, хто підписався нижче, уповноважена особа Пфайзер Менюфекчуринг Дойчленд ГмбХ, Мусвальдаль 1, 79108 Фрайбург Ім Брайсгау, Німеччина, підтверджую, що даний продукт був виготовлений для ринку України.

Ми гарантуємо якість цього продукту та те, що він був виготовлений відповідно до правил Належної Виробничої Практики, в результаті чого був випущений відповідний Сертифікат аналізу.

Цей сертифікат не може бути відтворений і призначений тільки для регуляторних органів України.

Цим засвідчую, що наведена вище інформація є аутентичною та достовірною. Ця партія продукту була виготовлена, включаючи стадії пакування/маркування та контролю якості, на приведеному вище виробництві (ах) в повній відповідності до вимог НВП місцевих регуляторних органів та із специфікаціями, зареєстрованими в країні призначення. Використана субстанція була виготовлена у відповідності до НВП. Технологічний процес, процес упаковки та документація з тестування були вивчені та визнані такими, що відповідають НВП. Можливі наслідки, відхилення або розбіжності було погоджено.

Ця серія була випущена Уповноваженою особою.

Цей сертифікат було створено валідованою системою та він є дійсним без ручного підпису.

Електронний підпис: Жаклін Шварц Місцева відмітка щодо дати випуску серії: 13 червня 2024 11:15:19

Позначка часу на сервері: 13 червня 2024 11:15:12

Документ розглядається як власність та був створений для операцій бізнесу та для перегляду робітниками та представниками регуляторних органів. Розповсюдження та передача третім сторонам без дозволу заборонено.

emaedcs060

Ідентифікаційний номер специфікації: 118441564

Ім'я LICRDE6P

	ПФАЙЗЕР МЕНЮФЕКЧУРИНГ ДОЙЧЛЕНД ГМБХ МУСВАЛЬДАЛЬ 1 79108 ФРАЙБУРГ ІМ БРАЙСГАУ НІМЕЧЧИНА	дата 02-09-2024 Сформовано: Наталія Вилензек Сторінка 4 з 4
---	---	--

Написання адреси та поштового індексу виробника ПФайзер Менюфекчуринг Дойчленд ГмбХ було змінено та раніше відомі як
 ПФайзер Менюфекчуринг Дойчленд ГмбХ
 Бетрібштетте Фрайбург
 Мусвальдаль 1
 79090 Фрайбург
 Виправлення поштового індексу та написання адреси виробника ПФайзер Менюфекчуринг Дойчленд ГмбХ не впливає на
 місцезнаходження виробничої ділянки та не впливає на вміст виробничої ліцензії та сертифікату НВП.



PFIZER MANUFACTURING
DEUTSCHLAND GMBH
BETRIEBSSTAETTE FREIBURG
MOOSWALDALLEE 1
79090 FREIBURG

Generated By:
Barbara Ciblis
Page 1 of 5

Certificate of Analysis

LIPRIMAR 20MG SMT 3X10 BLS UA

Destination: Ukraine

Material No.: F000142816

Batch No.: LG4956

Manufacturing Date: 15-NOV-2023

Expiry Date: 01-NOV-2026

Analytical Procedure: 966000-09

Test	Method	Limit	Result
APPEARANCE			
Result	I 2.02 / VISUAL	White, round film-coated tablets debossed "20" on one side and "ATV" on the other side.	conforms
DISSOLUTION			
ATORVASTATIN in 15 min			
Result	TM-0258A / UV	NLT 80% (Q) of the label claim	conforms
Average	TM-0258A / UV	-	92
Min.	TM-0258A / UV	-	87
Max.	TM-0258A / UV	-	100
Standard Deviation	TM-0258A / UV	-	4.7
Stage	TM-0258A / UV	-	1
IDENTIFICATION			
ATORVASTATIN / TLC			
Result	TM-0253A / TLC	The R(f) value for the principal spot obtained from the test solution corresponds to that obtained from the standard solution.	conforms
IDENTIFICATION			
ATORVASTATIN / HPLC			
Result	TM-0255A / HPLC	The retention time of the main peak obtained in the chromatogram of the assay preparation corresponds to that obtained in the chromatogram of the standard preparation.	conforms
PURITY			
ATORVASTATIN / HPLC			
INDIVIDUAL KNOWN DEGRADATION PRODUCTS			
PD140728			

Documentation is considered PROPRIETARY and is made available for business operations and review by employees and regulatory agencies. Distribution to third parties without prior permission is prohibited.

Certificate of Analysis

LIPRIMAR 20MG SMT 3X10 BLS UA

Destination: Ukraine

Material No.: F000142816

Batch No.: LG4956

Manufacturing Date: 15-NOV-2023

Expiry Date: 01-NOV-2026

Analytical Procedure: 966000-09

Test	Method	Limit	Result
Result	TM-0255A / HPLC	NMT 0.3% (Area)	0.2
PD130694 Result	TM-0255A / HPLC	NMT 0.3% (Area)	< 0.1
PD153824 Result	TM-0255A / HPLC	NMT 0.3% (Area)	<= 0.3
PD148996 Result	TM-0255A / HPLC	NMT 0.3% (Area)	<= 0.3
PD139884 Result	TM-0255A / HPLC	NMT 0.3% (Area)	<= 0.3
INDIVIDUAL UNKNOWN DEGRADATION PRODUCTS			
Result	TM-0255A / HPLC	NMT 0.2% (Area)	<= 0.2
TOTAL DEGRADATION PRODUCTS			
Result	TM-0255A / HPLC	NMT 1.5% (Area)	0.2
ASSAY			
ATORVASTATIN FCT / HPLC			
Result	TM-0255A / HPLC	19.4 - 21.0 mg (decl. 20 mg)	20.4
ASSAY			
ATORVASTATIN FCT / HPLC			
Result	TM-0255A / HPLC	97.0 - 105.0 %	102.2
UNIFORMITY OF DOSAGE UNITS			
/ UV			
Result	TM-0256A / PH. Conforms to Ph. Eur. EUR. 2.9.40 / 2.9.40 requirements USP-NF <905>		conforms
Average	TM-0256A / PH. - EUR. 2.9.40 / USP-NF <905>		101.7
Min.	TM-0256A / PH. - EUR. 2.9.40 / USP-NF <905>		100.1

Documentation is considered PROPRIETARY and is made available for business operations and review by employees and regulatory agencies. Distribution to third parties without prior permission is prohibited.

	PFIZER MANUFACTURING DEUTSCHLAND GMBH BETRIEBSSTAETTE FREIBURG MOOSWALDALLEE 1 79090 FREIBURG	Generated By: Barbara Ciblis Page 3 of 5
---	--	---

Certificate of Analysis

LIPRIMAR 20MG SMT 3X10 BLS UA

Destination: Ukraine

Material No.: F000142816

Batch No.: LG4956

Manufacturing Date: 15-NOV-2023

Expiry Date: 01-NOV-2026

Analytical Procedure: 966000-09

Test	Method	Limit	Result
Max.	TM-0256A / PH. - EUR. 2.9.40 / USP-NF <905>	-	103.2
Stage	TM-0256A / PH. - EUR. 2.9.40 / USP-NF <905>	-	1
Acceptance Value	TM-0256A / PH. Not more than 15% EUR. 2.9.40 / USP-NF <905>	-	3
Coefficient of Variation	TM-0256A / PH. - EUR. 2.9.40 / USP-NF <905>	-	1.0

The results on the Certificate of Analysis were transcribed upon Head of Quality Control approval.



PFIZER MANUFACTURING
DEUTSCHLAND GMBH
BETRIEBSSTAETTE FREIBURG
MOOSWALDALLEE 1
79090 FREIBURG

Generated By:
Barbara Ciblis
Page 4 of 5

Certificate of Compliance

LIPRIMAR 20MG SMT 3X10 BLS UA

Destination: Ukraine

Material No.:	F000142816
Batch No.:	LG4956
Manufacturing Date:	15-NOV-2023
Expiry Date:	01-NOV-2026
Yield:	22544 EA
Date of Release:	13-JUN-2024
Source Lots:	3193512
Registration No.:	UA/2377/01/01
API Name:	Atorvastatin Calcium
Package Type:	Blister

Dosage form: film-coated tablet

Name, address and number of manufacturing authorization license of manufacturing and quality control site:

PFIZER MANUFACTURING DEUTSCHLAND GMBH BETRIEBSSTAETTE FREIBURG, MOOSWALDALLEE 1,
79090 FREIBURG, Germany;
DE_BW_01_MIA_2024_0007

Microbiological testing: total aerobic microbial count (TAMC), method: Ph. Eur. 2.6.12, limit: not more than 10^3 CFU/ g; total combined yeasts moulds account (TYMC), method: Ph. Eur. 2.6.12, limit: not more than 10^2 CFU/ g; specified microorganisms, absence of E. coli (1 g), method: Ph. Eur. 2.6.13, limit: conforms.

This testing is not performed routinely, meets when tested. If no results are listed on the CoA, testing was not performed.

Certificate of Quality

To the Health Authorities of Ukraine

GUARANTEE LETTER

I, the undersigned, qualified person for the PFIZER MANUFACTURING DEUTSCHLAND GMBH BETRIEBSSTAETTE FREIBURG, MOOSWALDALLEE 1, 79090 FREIBURG, Germany confirm that this product was manufactured for the Ukrainian market.

We guarantee the quality of this product was manufactured according to the rules of Good Manufacturing Practice and the appropriate Certificate of Analysis was issued.

This certificate may not be reproduced and is solely destined to the Governmental Authorities of Ukraine.

Documentation is considered PROPRIETARY and is made available for business operations and review by employees and regulatory agencies. Distribution to third parties without prior permission is prohibited.



PFIZER MANUFACTURING
DEUTSCHLAND GMBH
BETRIEBSSTAETTE FREIBURG
MOOSWALDALLEE 1
79090 FREIBURG

Generated By:
Barbara Cihlis
Page 5 of 5

I hereby certify that all the manufacturing stages, including packaging / labeling and quality control, of this batch of finished product have been carried out in full compliance with the GMP requirements of the EU and with the requirements of the Marketing Authorization(s) of the destination country/countries. The above information is authentic and accurate. The corresponding active ingredient was manufactured in compliance with GMP. The batch processing, packaging and analysis records were reviewed and found to be in compliance with GMP. Any investigations, deviations and discrepancies have been approved.

This batch has been released by a Qualified Person.

The spelling and ZIP code of the manufacturer Pfizer Manufacturing Deutschland GmbH was changed to actual valid MIA: DE_BW_01_MIA_2024_0007. Formerly known as
Pfizer Manufacturing Deutschland GmbH

Betriebsstaette Freiburg

Mooswaldallee 1

79090 Freiburg

The correction of the ZIP code and spelling of the address of the manufacturer Pfizer Manufacturing Deutschland GmbH does not affect the localization of the manufacturing site or have any implications on the content of the manufacturing authorization and GMP certificate.

This CoA/CoC is an updated version (03) to a previously issued version. The following changes have been made:
Adress change in Header and Certificate of Compliance from current:

PFIZER MANUFACTURING DEUTSCHLAND GMBH MOOSWALDALLEE 1 79108 FREIBURG IM BREISGAU
GERMANY

to former:

PFIZER MANUFACTURING DEUTSCHLAND GMBH BETRIEBSSTAETTE FREIBURG MOOSWALDALLEE 1
79090 FREIBURG

Reference of current ZIP code was updated.

Actual valid MIA was added.

The product release date has not been changed.

Anna Victoria Welte

Anna Victoria Welte

09 Sep 2024 07:45:037-0400

REASON: I have reviewed this document and have no comments that
require addressing.

Data verified by

b9cd4ba2-0d14-41c6-a83a-06ce5efc935c

Manager SC / Date

Jacqueline Schwarz

Jacqueline Schwarz

09 Sep 2024 08:50:038-0400

REASON: I approve this document.

Data approved by

6dd619c9-9107-4cd0-add4-89c0d3de23e1

Qualified Person / Date

Documentation is considered PROPRIETARY and is made available for business operations and review by employees and regulatory agencies. Distribution to third parties without prior permission is prohibited.



Validation Report

1

Subject CN TAIGLE LLC
Subject DN EMAILADDRESS=operations@msbdocs.com,CN=TAIGLE LLC,OU=MSB,O=TAIGLE LLC,L=Irvine,ST=California,C=US
Email operations@msbdocs.com
Serial # 94498063721598734909473981570044396779
Issuer DN CN=Entrust Class 3 Client CA - SHA256,OU=(c) 2015 Entrust, Inc. - for authorized use only,OU=See www.entrust.net/legal-terms,O=Entrust, Inc.,C=US
Signing Time 09 Sep 2024 08:50:038-0400

The Certificate chain was successfully built to a Trusted Root Certificate.

The Signer's identity is valid.

The Document has not been modified since the signature was applied.

This page was added to the original document by the MSB Validation Service (SVS) as part of the process to convert active signature block content into inactive text. A full validation report of each signature is generated and cross-referenced using a numeric footnote annotation.



Audit Trail Report

Time Stamp	User	Document	Action	Details
09 Sep 2024 07:34:053-0400	Barbara Ciblis UUID : 752676f4-0aa0-4786-87b1-bd9aaddc448 Email : Barbara.Ciblis@pfizer.com IP Address : 168.224.160.14 OS: Windows, Browser: Edge, Device: Desktop		Started	The custodian composed the ePak successfully. Subject: LIPRIMAR 20MG SMT 3X10 BLS_LG4956_F000142816_EU_TR_MOW_CoA LIMS_Version 03_MSB ePak UUID: f339768e-76a8-4093-8c19-0546b7df199
09 Sep 2024 07:34:054-0400	Anna Victoria Welte UUID : b9cd4ba2-0d14-41c6-a83a-06ce5efc935c Email : AnnaVictoria.Welte@pfizer.com		RequestSent	Sign request sent to ePak recipient.
09 Sep 2024 07:44:033-0400	Anna Victoria Welte UUID : b9cd4ba2-0d14-41c6-a83a-06ce5efc935c Email : AnnaVictoria.Welte@pfizer.com IP Address : 168.224.160.14 OS: Windows, Browser: Edge, Device: Desktop	Name : LIPRIMAR 20MG SMT 3X10 BLS_LG4956_F000142816_ EU_TR_MOW_CoA LIMS_Version 03_MSB.pdf UUID : 9f2dec3-e402-40a5-bbbf-1762c02711f8 Document Source : GNOSIS	DocumentViewed	Document viewed by signer.
09 Sep 2024 07:45:037-0400	Anna Victoria Welte UUID : b9cd4ba2-0d14-41c6-a83a-06ce5efc935c Email : AnnaVictoria.Welte@pfizer.com IP Address : 168.224.160.14 OS: Windows, Browser: Edge, Device: Desktop	Name : LIPRIMAR 20MG SMT 3X10 BLS_LG4956_F000142816_ EU_TR_MOW_CoA LIMS_Version 03_MSB.pdf UUID : 9f2dec3-e402-40a5-bbbf-1762c02711f8 Document Source : GNOSIS	Signed	The recipient signed the document after authentication via login password. Comments: None Signing Reason: I have reviewed this document and have no comments that require addressing. Consent: I understand that my Electronic Signature is Equivalent to my Handwritten Signature and is therefore legally binding. My Electronic Signature will remain unique to me, and under no circumstance I am allowed to disclose my password to any individual which may allow unauthorized access to system. I understand that I am accountable and responsible for all actions associated with my Electronic Signature.
09 Sep 2024 07:45:037-0400	Barbara Ciblis UUID : 752676f4-0aa0-4786-87b1-bd9aaddc448 Email : Barbara.Ciblis@pfizer.com		EPakStateChanged	The ePak progressed to the next workflow state .
09 Sep 2024 07:45:037-0400	Jacqueline Schwarz UUID : 6dd619c9-9107-4cd0-add4-89c0d3de23e1 Email : Jacqueline.Schwarz@pfizer.com		RequestSent	Sign request sent to ePak recipient.
09 Sep 2024 08:50:011-0400	Jacqueline Schwarz UUID : 6dd619c9-9107-4cd0-add4-89c0d3de23e1 Email : Jacqueline.Schwarz@pfizer.com IP Address : 168.224.160.14 OS: Windows, Browser: Edge, Device: Desktop	Name : LIPRIMAR 20MG SMT 3X10 BLS_LG4956_F000142816_ EU_TR_MOW_CoA LIMS_Version 03_MSB.pdf UUID : 9f2dec3-e402-40a5-bbbf-1762c02711f8 Document Source : GNOSIS	DocumentViewed	Document viewed by signer.



Audit Trail Report

Time Stamp	User	Document	Action	Details
09 Sep 2024 08:50:038-0400	Jacqueline Schwarz UUID : 6dd619c9-9107-4cd0-add4-89c0d3de23e1 Email : Jacqueline.Schwarz@pfizer.com IP Address : 168.224.160.14 OS: Windows, Browser: Edge, Device: Desktop	Name : LIPRIMAR 20MG SMT 3X10 BLS_LG4956_F000142816_ EU_TR_MOW_CoA LIMS_Version 03_MSB.pdf UUID : 9f2dec3-e402-40a5-bbbf-1762c02711f8 Document Source : GNOSIS	Signed	The recipient signed the document after authentication via login password. Comments: None Signing Reason: I approve this document. Consent: I understand that my Electronic Signature is Equivalent to my Handwritten Signature and is therefore legally binding. My Electronic Signature will remain unique to me, and under no circumstance I am allowed to disclose my password to any individual which may allow unauthorized access to system. I understand that I am accountable and responsible for all actions associated with my Electronic Signature.
09 Sep 2024 08:50:038-0400	Barbara Ciblis UUID : 752676f4-0aa0-4786-87b1-bd9aaddc448 Email : Barbara.Ciblis@pfizer.com		Completed	The ePak is completed successfully.